

Lodz, 23-11-2021

**Certificate of Analysis No K/370/02/2021 (1/1)**

**Subject of analysis: UV-C STERILON MAX 1500 288W Philips 8x36W UVC radiators**

State of the subject: correct

**Customer: Lena Lighting S.A**  
**63-000 Środa Wlkp., ul. Kórnicka 52**

The device for testing was delivered by the Customer 03-11-2021  
The tests began: 16-11-2021  
The tests finished: 22-11-2021

Type of analysis	Analytical method	Results	
<b>Microbial parameters</b>			
Testing of the level of air pollution during the operation of the purifier in a room of 40 m <sup>2</sup>	Own methodology using a microbiological air sampler MAS-100 ECO™ Manual MAS-100 ECO™	*[cfu/1 m <sup>3</sup> ]	Microorganisms reduction
- Total Viable Count at time 0		<b>224</b>	-
- Total Viable Count after 2 hours		<b>64</b>	R <sub>2h</sub> = 71.43%
- Total Viable Count after 6 hours		<b>30</b>	R <sub>6h</sub> = 86.61 %
- Total Viable Count after 20 hours		<b>3</b>	R <sub>20h</sub> = 98.66%
-Total Yeast and Mold Counts at time 0		<b>166</b>	-
- Total Yeast and Mold Counts after 2 hours		<b>57</b>	R <sub>2h</sub> = 65.66%
- Total Yeast and Mold Counts after 6 hours		<b>25</b>	R <sub>6h</sub> = 84.94 %
- Total Yeast and Mold Counts after 20 hours		<b>5</b>	R <sub>20h</sub> = 96.99 %

\* The results are the average number of microorganisms from two measurements

Authorized:

Accepted:

KIEROWNIK  
Pracowni Mikrobiologii  
  
dr inż. Anna Szosland-Faltny  
Adiunkt

KIEROWNIK ZAKŁADU  
JAKOŚCI ŻYWNOSCI  
  
dr Beata Elżbieta

## Assessment of efficacy of UV-C STERILON MAX 1500 288W Philips 8 x 36W UVC radiators

### The aim and scope of the research

The aim of the study was to determine the effectiveness of air disinfection using **UV-C STERILON MAX 1500 288 W Philips 8x36 W UVC radiators** (Certificate of Analysis No K/370/02/2021) on the basis of reduction in numbers of molds, yeasts and bacteria that are present naturally in air, using aspiration method after 2, 6 and 20 hours of lamp working in a room with an area of 40 m<sup>2</sup>.

### Test procedure

The studies were conducted in accordance with the laboratory's methodology and the manufacturer's manual MAS-100 ECO<sup>TM</sup> (Microbiological Air Sampler) in a room with an area of 40 m<sup>2</sup>. Before turning on the device, the total viable count of microorganisms and the number of mold and yeast in the room air were examined (at 0 time). The bactericidal lamp was placed in the center of the room. The air pollution was measured after 2, 6 and 20 hours of operation. The tests were carried out using the aspiration method using the microbiological air sampler MAS-100 ECO<sup>TM</sup>. Each time the device took 1000 liters of air through a perforated plate (suction time about 9 minutes). The air stream containing particles was directed to the PCA or YGC agar surface in a standard Petri dish. After completing the air sampling cycle, the Petri dishes were incubated at 30°C for 72h or 25°C for 5 days, then the colonies grown were counted and the number of microorganisms in 1 m<sup>3</sup> of air was determined, taking into account the correction of the Feller's statistical correction table.

KIEROWNIK  
Pracowni Mikrobiologii  
  
dr inż. Anna Szosland-Faltyn  
Adiunkt